Supporting Information

Point-of-care human milk testing for maternal secretor status

Author names and affiliation: Saeromi Chung^a, Lars Bode^b, Drew A. Hall^{a,c,*}

^a Department of Electrical and Computer Engineering, University of California – San Diego, La Jolla, CA 92093, USA

^b Department of Pediatrics and Mother-Milk-Infant Center of Research Excellence (MOMI CORE), University of California – San Diego, La Jolla, CA 92093, USA

° Department of Bioengineering, University of California – San Diego, La Jolla, CA 92093, USA

* Corresponding author: drewhall@ucsd.edu



2'FL, and C. pH with 2.0µM 2'FL in 0.1 M PBS containing 5 mM Fe(CN)6^{3-/4-}.

Sample ID	EIS	HPLC (IIM)	Secretor/non-	Prediction
A	< <u>(µ1)</u> <330	<u>(μινι)</u> 5	Non-secretor	Correct
B	10129	11985	Secretor	Correct
С	7268	8197	Secretor	Correct
D	5323	4017	Secretor	Correct
Е	10243	11378	Secretor	Correct
F	<330	71	Non-secretor	Correct
G	8660	9159	Secretor	Correct
Н	4894	4632	Secretor	Correct
Ι	<330	34	Non-secretor	Correct
J	6126	7045	Secretor	Correct
K	<330	5	Non-secretor	Correct
L	9484	9585	Secretor	Correct

Table S1. Comparison of concentration of banked samples (HPLC vs. EIS)

Table S2. Cost analysis of proposed assay for 2'FL detection. All prices are from the published price on the vendor's website in low quantity.

Item		Unit Cost (USD)	Cost (USD)/batch (20×)
	Cys	\$15/10g	\$0.15
	UEA	\$44.40/1mg	\$0.44
Top 5 reagents	Glutaraldehyde	\$20/10mL	\$0.10
	10% BSA	\$24.20/10mL	\$0.01
	NH4OH	\$40/1L	\$0.10
Electrode	Au electrode	\$350/75	\$50.00

Cost per test